

File Number **18/31704758** English translation from the original in Spanish

## TEST REPORT

### Degrees of Protection Provided by Enclosures (IP Code)



#### Petitioner's Reference: SENSOCAR, S.A.U.

Customer Address: C/ Geminis, 77  
08228 Terrassa (Barcelona)

#### Equipment: Load cell

Brand:	SENSOCAR	Model:	SP-A
s/n:	76261/05.18	Internal identification number:	6502/1

#### Result:

The equipment described before has been tested according to the applied specifications\* and has been found in conformity with a degree of protection:

**Degree of protection (IP): IP6X**

\* See applied specifications in section 1.4.

#### Applied Standards:

UNE-EN 60529:2018 +/A1:2018 +/A2:2018 (EN 60529:1991 +/A1:2000 +/A2:2013)

**Date of issue:** Bellaterra, June 29, 2018

Albert Marginet Morales  
Technical Manager  
Electrical and Electronics  
LGAI Technological Center S.A.

The results refer only and exclusively to the sample, product or material delivered for testing in "Equipment" section above. The equipment has been tested under conditions stipulated by standard(s) quoted in this document.

This document will not be reproduced otherwise than in full.

This is the first page of the document, which consists of 10 pages of which 5 are annexes.

In case of discrepancy between English and Spanish version, the Spanish version prevails over the English one.